



ULTRA LEVEL TECH.CO.,LTD.

4F., No. 2, Lane 235, Baoqiao Road, Xindian Dist., New Taipei City 231, Taiwan

TEL:+886-2-2918-0118

FAX:+886-2-2918-1465

Subject:	Product Change Notice (PCN)	Date:	2017/9/12
-----------------	------------------------------------	--------------	-----------

Dear Customers:

ULTRA LEVEL is dedicated to providing the top quality and reliable products for our valued customers. To ensure that our product can keep meeting market demand, adopting forward-looking technology and meet the latest safety requirement, this notice is to inform you of a product change.

Products affected:

PRL1103x-y and PRL1103Dx-y (x=U, E or F; y=12, 15, 24, 36, 48, 0512 or 1224) series


Reasons for change:

Input terminal block connector change for IEC 60601-1: 2005+ A1, EN60601-1: 2006+A11+A1+A12 requirement.

Description of change:

IEC 60601-1: 2005+ A1, EN60601-1: 2006+A11+A1+A12 requires input connector needs to have European approval (TUV, VDE or EN/EC etc..). **New connector does not have plastic cover.**


For U & F type - Current terminal block-



New terminal block-



For E type - Current terminal block-



New terminal block-



All shipment after Jan, 1, 2018 will be with new terminal block.

Sincerely,

Sam Tu

President / Ultra Level Tech. Co., Ltd.