



# ULTRA LEVEL TECH. CO., LTD.

4F., No. 2, Lane 235, Baoqiao Road, Xindian Dist., New Taipei City 231, Taiwan  
TEL:+886-2-2918-0118 FAX:+886-2-2918-1465

<b>Subject:</b>	<b>Product Change Notice (PCN)</b>	<b>Date:</b>	2017/4/11
-----------------	------------------------------------	--------------	-----------

**Dear Customers:**

ULTRA LEVEL is dedicated to providing the top quality and reliable products for our valued customers. To ensure that our product can keep meeting market demand and adopting forward-looking technology, this notice is to inform you of a product change.

**Products affected:**

RL0402E-x and RL0402DE-x x=03, 05, 12, 18, 24, 36, 48, 0312, 0324, 0512, 0524 or 1224 Medical series

**Reasons for change:**

Input terminal block connector change for IEC 60601-1: 2005+ A1, EN60601-1: 2006+A11+A1+A12 requirement.

**Description of change:**

IEC 60601-1: 2005+ A1, EN60601-1: 2006+A11+A1+A12 requires input connector needs to have European approval (TUV, VDE or EN/EC etc..). New connector does not have plastic cover as below picture.

Current terminal block-



New terminal block-



All shipment after Jun 30, 2017 will be with new terminal block.

Sincerely,

Fiona Wu  
General Manager / Ultra Level Tech. Co., Ltd.